# **Di-iso-PROPYL AMINE**



## 1. PERFORMANCE

1) Measuring range 1-16 ppmNumber of pump strokes  $1(100 \text{m} \ell)$ 

2) Sampling time : 1 minute/1 pump stroke

3) Detectable limit : -4) Shelf life : 3 years 5) Operating temperature :  $15 \sim 25 \,^{\circ}$ C

6) Reading : Graduations printed on the tube are calibrated by Ammonia at 1 pump stroke

and Di-iso-Propyl amine concentration is determined by using a conversion

chart at 1 pump stroke.

7) Colour change : Pale purple → Pale Yellow

#### 2. RELATIVE STANDARD DEVIATION

RSD-low: 10% RSD-mid.: 5% RSD-high: 5%

### 3. CHEMICAL REACTION

By reacting with Phosphoric acid, PH indicator is discoloured. [(CH₃)₂CH]₂NH + H₃PO₄→ (R₂NH₂)₃PO₄

#### 4. CALIBRATION OF THE TUBE

PERMEATION TUBE METHOD

#### 5. INTERFERENCE AND CROSS SENSITIVITY

Substance	Interference	Coexistence
Amines	Similar stain is produced.	Higher readings are given.

