
# **FORMALDEHYDE**



## 1. PERFORMANCE

1) Measuring range 1-35 ppm Number of pump strokes  $3(300 \text{m} \ell)$ 

2) Sampling time : 3 minutes/3 pump strokes

3) Detectable limit 0.5 ppm4) Shelf life 3 years5) Operating temperature  $0 \sim 40 \text{ °C}$ 

6) Reading : Direct reading from the scale calibrated by 3 pump strokes

7) Colour change : White→Brownish orange

## 2. RELATIVE STANDARD DEVIATION

RSD-low: 15% RSD-mid.: 15% RSD-high: 15%

### 3. CHEMICAL REACTION

By reacting with Aromatic compounds, a polymer is produced.

$$HCHO + \bigcirc + H_2SO_4 \longrightarrow \left[ \bigcirc \bigcap_{H}^{C} \bigcirc \bigcap_{n} + H_2O \right]$$

## 4. CALIBRATION OF THE TUBE

ABSORPTIOMETRIC METHOD

### 5. INTERFERENCE AND CROSS SENSITIVITY

Substance	Interference	ppm	Coexistence
Styrene	Similar stain is produced.		Higher readings are given.
Trichloroethylene		500	"
Ethyl acetate		1,000	"
Ether		1,000	"
Acetaldehyde	Similar stain is produced.	1	"